

Computer failure

Danger: You are blind, don't know what is going on.

1. Common occurrence.

2. Reboot IOC

1. Ask MCC for beam off

2. Use manual HPH PS & analog temperature readout to control tgt temp during reboot

3. Eventually we should have manual JT control but at present this important feature is missing

1. May need to enter Hall C to close JTs

3. Should happen much less frequently now

1. Our IOC is now in the CH, not in Hall C

