

High Loop Pressure Alarm

Danger: Bursting cell window

1. Diurnal pressure variation?

1. Check archiver to see recent history.

2. Temps high? Turn off the beam.

1. System is warming up.
2. Warm (storage) pressure ~ 66 psia.
3. See High temp alarm procedures:

1. Probably the LH2 pump has tripped, or else a vacuum pump has tripped.
2. Call expert if problem or solution not obvious.

