

## Hall D Controls Meeting 15-Nov-2018

**Date: 11/15/18**

**Time: 10:00am**

*Attendees: Pablo Campero, Hovanes Egiyan, Brian Eng, Amanda Hoebel, Vanik Kakoyan, Tyler Lemon, Nick Sandoval, Scot Spiegel, Tim Whitlatch*

### 1. General:

- 1.1. Downtime scheduled for Nov 26-29.
  - 1.1.1. Solenoid ramped down and stay down during downtime.
  - 1.1.2. ComCal commissioning during downtime.

### 2. Solenoid status:

- 2.1. Liquid level testing planned during downtime.

### 3. ComCal:

- 3.1. Planned commissioning during downtime.
- 3.2. Environment monitoring variables archived.
  - 3.2.1. Fan speeds not archived.
- 3.3. Interlocks checked.
- 3.4. Control board for LED pulser installed.
- 3.5. Voltages need to be added into GUI.
- 3.6. Resetting voltage crate after interlock investigated.
  - 3.6.1. PLC and voltage crate latch during interlock.
  - 3.6.2. Voltage crate latch may be disabled and use only PLC latch.
- 3.7. Chiller will be installed on Nov 26.
  - 3.7.1. Will be interlocked.
- 3.8. GUI made for VME scalars.
  - 3.8.1. Translation table is not correct.