

cRIO Test Stand Status Report
July 3, 2019

1. Wrote manual voltage tests for NI 9219 for various ranges.
 - 1.1. Tests: differential nonlinearity, dynamic range, gain error, integral nonlinearity, and offset error
 - 1.2. Voltage ranges: ± 60 V, ± 15 V, and ± 4 V.