cRIO Test Stand Status Report July 10, 2019

- 1. Completed manual voltage tests for NI 9219.
 - 1.1. Tests: differential nonlinearity, dynamic range, gain error, integral nonlinearity, and offset error
 - 1.2. Voltage ranges: ± 1 V, and ± 0.125 V.
- 2. Wrote and tested one of five automatic tests ($\pm 60 \text{ V}$ range).