



Detector Support Group

Weekly Report, 2019-02-06

Summary

Hall C EPICS

- Python EPICS channel access module (“pyepics”) investigated for use in alarm handling and as a backup and restore tool.
 - * Module has built-in auto-save and restore of PVs, but does not seem to consistently be able to record PV field values.
 - * Module has alarm features that monitor PVs and then take action in Python if there is an alarm.
- *ProcServ* and telnet installed on DSG-C-LINUX1 PC.
 - * *ProcServ* is a program that allows IOCs to run as a background process.
 - * Telnet is used to access *procServ* process while it is running in background to be able to restart or kill process.
 - * Developed shell script to start and stop a test IOC.
- Conversion from HV Tcl/Tk monitoring and control system to HV CSS analyzed for HMS and SHMS HV alarm systems.
- EPICS Alarm Specifications and Alarm Severity investigated, alarms to be implemented in future alarm handler for Hall C EPICS.
 - * Four alarms types available: *Scan Alarms*, *read/write Alarms*, *Limits Alarms* and *State Alarms*
 - * Four Alarm Severities available: *No Alarm*, *Minor*, *Major* and *Invalid*.
 - * Some possible Alarm Status for analog value limits are: *High-High*, *Low-Low*, *High* and *Low*.
- Started DSG note on the *Hall C EPICS Slow Controls and Monitoring System Proposal*.

HDice

- Issue of increased NMR cycle time after the upgrade of computer to Windows 10 on rack 1 under investigation; considered items to check are:
 - * GPIB hardware setup parameters.
 - * Resources being used by system processes.
 - * Computer specifications (new computer purchased by HDice Group)

LTCC

- Gas System LabVIEW code modified to update set points via auto-updating configuration file started.
 - * Code will replace the hard coded values for each set point.
 - * Code will allow the set points to be updated as they are changed on the GUI.
- Cabling for the H₂O sensor and C4F₁₀ supply solenoid fabricated and installed along with a 120VAC coil.
- New LTCC EPICS signals added to Gas Shed cRIO which required a reboot.
 - * Signals added: Gas pressure, pump status, valve status and water ppm value.

Hall B Gas

- Still waiting on NI to respond about Forward Carriage cRIO reboot problems.
 - * Problem presented to NI Research and Development department.


RICH

- N₂ backup manifold installed and operational in Hall B.
 - * Backup manifold installed on N₂ panel with two 12-bottle N₂ banks acting as backup supply.
 - * Backup supply should last ~5 hours with RICH’s 40-slm N₂ flow rate.
 - * Two 0 – 50 psi absolute pressure transducers wired to RICH N₂ cRIO with Pablo and Amanda.

Detector Support Group

Weekly Report, 2019-02-06

- * PTs measure backup supply pressure (normally ~12 psi) and main N2 supply pressure (normally ~35 psi).
- * N2 cRIO LabVIEW program modified to convert PTs' absolute pressure reading to gauge pressure.
- * Added indicators for new PTs to N2 cRIO EPICS screen.
- * New PTs added to Mya.



Instrumentation for RICH N2 backup supply system installed

- Ways to improve $d\theta$ test investigated, test is used to measure spherical mirrors' radius of curvature.
 - * Possible improvements:
 - Development of one program that performs all actions for CCD stage movement, CCD data acquisition, image analysis, and mirror alignment.
 - Move of test program to a PC with network access.
 - Design, better mirror stand for test station.
 - * Source code for test's executables requested from INFN collaborators.

RTPC

- RTPC Detector Gas System Status created.
- Agenda for Controls meeting distributed, meeting to be held on 2/7/2019.

Accelerator Division

- Population of one of six VME FSD boards continued for Machine Protection System. Engineering Division.
 - * Second of six boards completed.
 - * 210 capacitors (1uF, 25V, 10%, 0603) soldered on three boards.



Detector Support Group

Weekly Report, 2019-02-06

cRIO test station

- Debugged “Unknown” LabVIEW error generated during the test of NI-9263 module.
 - ★ Error defined as “Unknown (0x00000000)” by NI made LabVIEW crashing with no option for the regular debugging process.
 - ★ Verified proper communication between cRIO controller and PC in Hall B Dev subnet; no issues found.
 - ★ Checked LabVIEW libraries used to send and receive messages between Real Time running in cRIO and User Interface running in the computer; running with no problem
 - ★ Found that the wiring between NI-9263 analog output module and Keithley 2002 Voltmeter was the cause to generate unknown LabVIEW error.
 - Internal wiring issue found in the module was making the LabVIEW crash every time that LabVIEW programs was setting a voltage output in the module.
 - ★ Problem and errors solved by wiring the NI 9263 outputs to rear card of the Keithley, so that it can read up to 8 channels at once, isolating each channel.
- NI 9263 module wired and made drawing of wiring.
- LabVIEW code written to add drop-down menu on User Interface front panel.
- LabVIEW SubVI code written to read samples and added to cRIO Test Station program.

PLC Test Station

- New IP address assigned for dsg-plc-135 PLC
 - ★ PLC is intended to be used in the DSG- CompactLogix PLC test station.
 - ★ Ran Ethernet cable and connected PLC to the Hall B Dev subnet.
- Connected power cable 110 VAC to the dsg-plc-135 PLC power supply.
- Installation of Rockwell Software on DSG-COMP2 PC is in progress.

DSG Website

- Directory structure on the shared “O” drive and the DSG website web links created to link files with directory structures.