

## Target Installation Status

Vincent Sulkosky

December 6, 2002

### Installation Status

- Target system and electronics moved to Hall A and Counting House
- Electronics set up in Counting House
- Helmholtz coils connected and calibrated
- Large and small coil field map done
- Local control works for the target lifting motor
  - Limit switches need adjustment (K. McCormick)
  - EPICS Control (K. McCormick)

### Short Term Plan

- Set H0 with coils at  $19^\circ$  from beamline
- Install target ladder for survey (Tuesday)
- RF field calibration