## Jefferson Lab Alignment Group DATA TRANSMITTAL

TO: J. P. Chen, J. LeRose DATE: 01 June 2001

FROM: Jim Dahlberg Checked: #: A672

## **DETAILS:**

Below are the results of the Hall A He3 target ladder and compass survey performed on May 10<sup>th</sup> and 24<sup>th</sup>, 2001. The numbers shown for the target ladder represent the amount the survey target would have to move (in mm) to be in the ideal location. A negative X is a movement to the beam right, positive Y is up, and positive Z is downstream. The compass measurements are the average of normal and reverse fields.

Target	Del Z	Del X	Del Y
ATGLADE	0.52	0.39	0.67
ATGLADF	-0.18	0.52	0.63
ATGLADG	-0.61	0.24	0.18
ATGLADH	0.54	0.38	0.25
ATGLADI	0.22	0.06	0.88
ATGLADJ	-0.79	0.19	0.76

Compass parallel to beam would have to rotate 0.03deg. ccw to be in ideal position. Compass perpendicular to beam would have to rotate 0.20 deg. cw to be at ideal position.

## TARGET LADDER FRAME

