



# Jefferson Lab Alignment Group

## Data Transmittal

**TO:** J. P. Chen, J. Benesch

**DATE :** Jul 29, 2003

**FROM:** J. Dahlberg

**Checked:**

**#** A881

### DETAILS:

Below are the results from the July 11, 2003 survey performed on the Hall A super harp/BPM girder. A positive X indicates the center is to the beam left from ideal and a positive Y is above. The values are in millimeters.

	<b>X</b>	<b>Y</b>
IPM1H04A	-0.13	-0.25
IHA1H04A	0.13	-0.04
IPM1H04BH	0.10	-0.44
IPM1H04BV	0.10	-0.22
IPM1H04CH	0.06	-0.05
IPM1H04CV	-0.10	0.10
ITV1H04A	-0.08	-0.03
IPM1H04D	0.35	0.01
IHA1H04B	-0.03	0.02