



# Jefferson Lab Alignment Group

## Data Transmittal

**TO:** Melvin Washington, John Jefferson

**DATE:** 06 Nov 2019

**FROM:** Robert Mares

**Checked:** LEB

**# :** M1946

**DETAILS:**

M:\align\DATA\Inspection\RadCon\RAD CON CALIBRATION BENCH

On November 5, 2019, a survey was conducted on the Radiation Control calibration table to confirm the accuracy of distances displayed by the controls were in accordance with physical distances between the source and target. Survey results are as follows:

Controller Display: 1.0000 meter

Physical Distance (as measured): 0.9997 meter