



Jefferson Lab Alignment Group

Data Transmittal

TO: P.King,T.Horn,C.Munoz Camacho,S.Malace,J.Butler

DATE: 17 Apr 2018

FROM: Steve Hardisty

Checked: TPS III

: A1859

DETAILS:

data : aalign\electron\2018\H041718A

Below are the results from the survey of the left spectrometer on April 17, 2018. The horizontal pointing value shows how much the central axis of the spectrometer misses the ideal target. This value is perpendicular to the spectrometer axis, not along the beam line. For the vertical pointing, a positive value indicates that the spectrometer is pointing above the target.

A graphical sketch is shown at the end of this transmittal

===== RESULTS ===== E041718A

The central ray of the spectrometer is at -20.880 degrees
The central ray is missing the defined target center by 1.66 [mm] Downstream
and -0.69 mm vertically [positive value is up]

If the offset is corrected by secondary alignment, the spectrometer will be at -20.892 degrees

To achieve this optimal setting
make the following adjustments:
spectrometer will be at -20.892 degrees
Horizontal corrections:

Move rear jacks along tangent -1.69 mm Downstream

9 Par A posteriori value : 0.13 (mm)

E041718A

Beam Direction

- Beam-Spec Intercept Point
- Beam-Spec Perpendicular Point
- Spectrometer Projected Target Point
- ✕ Straight-Ahead Target Point [ideal]

angles: delta : 20.88035 [degrees]
beam : 142.49998 [degrees]
spectrometer: 163.38033 [degrees]

perpendicular distance : 1.656 [mm]
target - intersect dis : 4.646 [mm]
found target - intersect dis : 5.798 [mm]
Spectrometer is -0.69 lower than ideal target [mm]

- Spectrometer Line
- Straight-Ahead Beam
- Perpendicular line

