



Jefferson Lab Alignment Group

Data Transmittal

TO: E. Abkemeier, A. Saha, M. Spata, S. Suhring

DATE: Dec 21 2004

FROM: Richard Schwartz

Checked:

: A967

DETAILS:

Below are the results of the Hall A dump survey carried out on 21st December, 2004. The values refer to the component's location relative to its ideal position. These values are in millimeters with BL signifying Beam Left, and positive Y being high. The X and Y of the aperture and the Y for the viewer is estimated to be accurate to +/- 1.0 mm. The X position of the viewer is accurate to an estimated +/- 3.0 mm.

As-Found Position

	X	Y
0.50" aperture	1.9 BR	-1.9
Viewer Target	13.3 BL	-0.7