



Quartz Bar Hodoscope - 32 x ET9814WB PMTs Photon Detection Efficiency Validation



ET9814WB - Photon Detection Efficiency Measurement

- Vendor ET provided complete scan in UV-VIS (200-700 nm) for all PMTs
- As alternative, verified performance by using *pulsed method* to measure actual *photon detection efficiency (PDE)* at 340 nm (as per purchase requisition spec)
- This will take into account *photoelectron collection efficiency* and *QE* of PMT
- Initial Specs in Purchase Order:
 - *At $\lambda=340\text{nm}$ - $QE \geq 25\%$ for all PMTs and at least 22 have $QE \geq 30\%$*
 - *All PMTs must have $\text{Gain} \geq 0.4 \times 10^7$ at $HV = 1.9 \text{ kV}$*

Measure Photodetection Efficiency (PDE)

$$PDE = \frac{N_{pe}^d}{N_{ph}}$$

Mean number of
detected photons

Mean number of
received photons

$$N_{pe}^d = \mu_{pe}^{100\%}$$


Mean photoelectron
count scaled to
100% (full) intensity

$$N_{ph} = I_{ph} \cdot A_{pmt}$$


Calibrated intensity
[photons / mm²]

Photosensitive
area of PMT
(Vendor spec)


PDE measurement setup




Thorlabs 340 nm LED






Photon intensity calibration



Hamamatsu S2281
100 mm² calibrated photodiode






Photon Intensity Calibration and intensity Controls



Photon Intensity Calibration and intensity Controls

Set Filter to 0.1% and LED level to 30%
Stepped through 20, 40, 60, 80 & 100%





Charge Spectra analyzed via Pavel Degtiarenko's Fit Program


Pavel Degtiarenko, "Precision analysis of the photomultiplier response to ultra low signals", Nuclear Instr. and Methods in Physics Research, A 872 (2017) 1-15

2017/10/04 ;

ET20294_100.txt




Collect Charge Spectra through 5 filter settings



Photon Detection Efficiency (PDE)





Mean Detected Photon Count



Use line fit from plot of *mean photon count* as function of *filter setting* to extrapolate to mean detected photon count at full intensity (filter setting=1)

Ratio of mean detected photons to actual photon intensity gives PDE

Results







BACKUP



Jefferson Lab

Internal luminescence activated when signal loss occurs



UV window / RTV615 gel pad transmission

Sample (#)	Mass (g)	Thickness (mm)	Curing Method (Temperature °C)
1	4.20 ± 0.01	1.90 ± 0.07 ± 0.02	Normal (25 °C)
4	7.31 ± 0.01	3.35 ± 0.07 ± 0.02	Normal (25 °C)
6	5.05 ± 0.01	2.30 ± 0.02 ± 0.02	Normal (25 °C)
7	3.20 ± 0.01	1.48 ± 0.05 ± 0.02	Heat (100 °C)
8	3.03 ± 0.01	1.37 ± 0.02 ± 0.02	Normal (25 °C)
9	7.37 ± 0.01	3.39 ± 0.02 ± 0.02	Heat (100 °C)

