

Current Trip Testing of CAEN High Voltage Modules - 740 μ A

#	Board #	Analysis (1/7/2021)	Status
1	128	channel 15 decreased, increased, then decreased to 0 μ A	
2	173		
3	184		
4	256		
5	262		
6	297	non-functional, unable to be tested	
7	299		
8	301	tested twice, ch 16 trip initiated at different levels	
9	302		
10	304		
11	309		
12	313		
13	314		
14	316		
15	317		
16	324		
17	325		
18	326		
19	327		
20	336		
21	337		
22	338		
23	339		
24	347		
25	349	channel 13 will not ramp above 40 μ A	
26	350		
27	351		
28	352		
29	353		
30	354		
31	355		
32	356		
33	358		
34	359		